



CRYOPRESERVED CARDIAC AND VASCULAR ALLOGRAFT BIO-IMPLANTS

INSTRUCTIONS FOR USE

Read this entire package insert carefully prior to use.
Federal law (USA) restricts this allograft bio-implant for use by a licensed clinician only.

DESCRIPTION

This allograft bio-implant was processed from donated human tissue, resulting from the generous gift of an individual or his/her family. Processing involves dissection, antibiotic disinfection and cryopreservation.

INDICATIONS FOR USE

This allograft bio-implant is intended for implantation.

CONTRAINDICATIONS

The contraindications include, but are not limited to:

- Use in any patient who has a known or suspected allergy to any of the antibiotics and/or processing reagents listed in this package insert.

WARNINGS AND PRECAUTIONS

The same medical/surgical conditions or complications that apply to any surgical procedure may occur during or following implantation. The surgeon is responsible for informing the patient of the risks associated with their treatment and the possibility of complications or adverse reactions. As with any allograft bio-implant, the potential for transmission of infectious agents exists.

This bio-implant may contain residuals of antibiotics (Ciprofloxacin, Gentamicin, Lincomycin, Mefoxitin, Meropenem, Polymyxin B Sulfate, and/or Vancomycin), Anidulafungin and dimethylsulfoxide (DMSO). The concentration of DMSO in the final packaging is approximately 10%. Following the thaw and dilution instructions in this insert will result in a residual DMSO of approximately 2%. Caution should be exercised if the patient has a known sensitivity to any of these antibiotics and/or reagents. This bio-implant and its packaging materials are latex-free.

STORAGE REQUIREMENTS

The distributor, intermediary and/or end-user clinician or facility is responsible for storing this allograft bio-implant under appropriate conditions prior to further distribution or implantation. Bio-implants must be stored as listed in the table below. Storage conditions are also stated on the label.

Preservation Method	Storage Temperature	Special Cautions
Cryopreserved	Store at liquid nitrogen vapor phase temperature (-120°C or below).	Store in its original cardboard box.
		Store in liquid nitrogen vapor phase freezers or equivalent, or maintain in the cryoshipper until removed for thawing and dilution.
		Do not submerge directly into liquid nitrogen.

! If a cryopreserved bio-implant is accidentally immersed in liquid nitrogen, contact LifeNet Health immediately. The circumstances surrounding accidental immersion must be reviewed prior to the potential use of the bio-implant.

DONOR SCREENING AND TESTING

All donors have been screened and tissues recovered, processed, stored, tested and distributed in accordance with current U.S. federal regulations as promulgated in 21 CFR 1270 and 1271, current Standards for Tissue Banking set forth by the American Association of Tissue Banks (AATB) and international laws and regulations as required. This allograft bio-implant was deemed suitable for implantation by LifeNet Health. A physician medical director evaluated the following donor variables to determine donor suitability: infectious disease test results, current donor medical history, behavioral risk assessment interview, physical assessment, relevant medical records, including previous medical history, laboratory test results, and autopsy or coroner reports (if performed).

All donors are tested for relevant infectious diseases. Testing is performed by laboratories that are registered with the U.S. Food and Drug Administration (FDA) and certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA) and 42 CFR 493. Test methods that are FDA-licensed, approved, or cleared for donor screening are used as available. The following test criteria were met for the donor of this allograft bio-implant:

Required Infectious Disease Testing	
Test	Acceptance Criteria
HBcAb: Hepatitis B Total Core Antibody	Negative/Non-Reactive
HBsAg: Hepatitis B Surface Antigen	Negative/Non-Reactive
HCV NAT: Hepatitis C Virus Nucleic Acid Test	Negative/Non-Reactive
HCVAb: Hepatitis C Antibody	Negative/Non-Reactive
HBV NAT: Hepatitis B Virus Nucleic Acid Test*	Negative/Non-Reactive
HIV-1 NAT: Human Immunodeficiency Virus Type 1 Nucleic Acid Test	Negative/Non-Reactive
HIV 1/2 Ab: Human Immunodeficiency Virus Types 1/2 Antibody	Negative/Non-Reactive
RPR/STS or Equivalent: Syphilis	Confirmatory Negative/Non-Reactive
HTLV I/II Ab: Human T-Lymphotropic Virus Types I/II Antibody**	Negative/Non-Reactive

* Not required for donors recovered prior to December 16, 2016. Performed as required by international laws and regulations.

**Not required for donors recovered after March 31, 2010. Performed as required by international laws and regulations.

POTENTIAL ADVERSE EVENTS

Potential adverse events or outcomes include, but are not limited to, infection, allograft tissue rejection, allergic reaction to residual processing reagents, fibrocalcification, stenosis, hemorrhage, thromboembolism, loss of bio-implant structural integrity, bio-implant rupture, aneurysm, reoperation and/or death. Promptly report any adverse event(s) or outcome(s) potentially attributable to the allograft bio-implant (See COMPLAINTS AND RETURNS section).

TRACEABILITY

It is the responsibility of the end-user to maintain recipient records for the purpose of tracking tissue post-implantation. As a courtesy to the end-user clinician or facility, LifeNet Health has enclosed a Graft Implant Tracking Card to assist in the post-implantation tracking. Please refer to the enclosed card for additional instructions.

COMPLAINTS AND RETURNS

For further information on returns or to report a complaint or adverse event, please contact your authorized distributor or LifeNet Health Client Services (available 24 hours a day) at 1-888-847-7831 (inside the U.S.) or 00+1-757-464-4761 ext. 2000 (outside the U.S.) and have the bio-implant's identification number available (see label).

Cryoshipper Return: Please refer to CardioGraft Tissue Receiving & Cryoshipper Return Instructions. This document accompanies each cryoshipper.

WARRANTY STATEMENT

Due to the inherent variability of allograft tissue, biological and biomechanical properties cannot be guaranteed by LifeNet Health.

Please See Back for Thawing and Dilution Instructions

Manufactured by: LifeNet Health, 1864 Concert Drive, Virginia Beach, Virginia, 23453 USA

Source Establishment: LifeNet Health CTO #100038

For patent information, please visit: www.lifenethhealth.org/patents

LifeNet Health and the LifeNet Health logo are registered trademarks of LifeNet Health.
 ©2025 LifeNet Health, Virginia Beach, VA. All rights reserved.

63-001713

INSTRUCTIONS FOR USE



The cardiovascular allograft is encased in a protective “CryoCushion” within the cryoshipper. When receiving multiple allografts, it is imperative that the cushion be immediately and carefully returned to the cryoshipper after the desired allograft is selected.

GENERAL INSTRUCTIONS

- Use on a single occasion for a single patient only.
- Once the packaging is opened, the bio-implant must be used for the current procedure or discarded.
- Inspect the bio-implant, inner and outer packaging, and labels carefully:
 - Do not use past the expiration date as indicated on the label.
 - Do not use if the bio-implant is damaged or the packaging integrity is compromised.
 - Do not use if there are discrepancies in label information.
- If using more than one bio-implant in the same procedure, ensure to maintain proper identification of each bio-implant type after it is removed from its outer packaging.
- Use aseptic technique at all times.
- Do not sterilize.
- Keep the bio-implant stored according to recommended storage instructions until preparing it for implantation.
- Thaw and dilute each bio-implant individually.
- Once the bio-implant is thawed, it must be used for the current procedure or discarded.
- Do not refreeze the bio-implant after thawing has begun.
- Dropping or jarring may compromise the integrity and/or functionality of the bio-implant.

PREPARATIONS FOR USE:

Supplies Needed:

- 2 major basins (≥ 5000 ml capacity)
- 1000 ml room-temperature D5LR sterile solution
- Non-toothed clamp
- 2000 ml warm 0.9% (normal) sterile saline
- Scissors
- 20 cc syringe (Vascular Bio-Implants Only)
- Smooth forceps
- Thermometer

Note: Additional warm or room temperature sterile saline may be required to maintain thawing temperature or utilize a sterile warmer that can be maintained at 37°C–42°C.



This thawing process helps minimize damage to the bio-implant by controlling the melting of ice crystals and considerably diluting the cryoprotectant, DMSO. Failure to follow these instructions could result in failure of the bio-implant and be detrimental to patient safety. Please coordinate the timing of this procedure with the surgeon.

OPENING AND THAWING INSTRUCTIONS:

(Cardiac & Vascular Bio-Implants)

1. **Non-Sterile Team Member:** Using insulated gloves, retrieve the packaged bio-implant from the appropriate storage location. Open the box lid and **place at room temperature for 7 minutes on a stable surface.** While the bio-implant is thawing, pour a minimum of 2000 ml warmed sterile saline solution into the first sterile major basin. Add enough room temperature sterile saline to maintain a temperature at 37°C – 42°C.
2. After the 7 minutes has elapsed, remove the bio-implant from its box, open the outer peel pack and present the inner pouch to the **Sterile Team Member.**
3. **Sterile Team Member:** Using a non-toothed clamp, firmly grasp the inner pouch by the sealed edge and remove from outer peel pack.
4. Slowly lower inner pouch into the pre-filled major basin. **The saline must maintain a temperature of 37°C to 42°C throughout the entire thawing process.** To monitor the temperature, use a sterile thermometer or a sterile warmer that can be set to a temperature of 37°C to 42°C. If using a thermometer, add warm or room temperature sterile saline as appropriate to maintain the required temperature range. Do not exceed 42°C. Exceeding 42°C may damage the bio-implant. Do not pour saline directly onto the pouch; pour into surrounding saline bath.
5. Gently agitate solution around pouch periodically for 5-7 minutes until ice is melted. Do not manipulate the bio-implant.
6. Once the bio-implant has thawed, dry the outside of the pouch thoroughly. Open with scissors.
7. Continue to the dilution instructions appropriate for the bio-implant type.

DILUTION AND INSPECTION INSTRUCTIONS:

FOR CARDIAC IMPLANTS:

(Cont. from Opening and Thawing Instructions)

8. **Non-Sterile Team Member:** Pour 1000 ml of room temperature D5LR into the second sterile major basin.
9. **Sterile Team Member:** Gently remove the bio-implant from the pouch with smooth forceps and carefully place into the basin containing the D5LR. Allow the bio-implant to passively rinse in this solution for a minimum of 5 minutes.

FOR PEDIATRIC CONDUIT IMPLANTS:

(Cont. from Opening and Thawing Instructions)

8. **Non-Sterile Team Member:** Pour 1000 ml of room temperature D5LR into the second sterile major basin.
9. **Sterile Team Member:** Gently remove the bio-implant from the pouch with smooth forceps and carefully place into the basin containing the D5LR. Remove the check valve at the proximal end of the vessel (with long suture tie) to allow perfusion.
10. Connect 20 cc syringe to the distal end (marked with a short suture tie) and gently flush the bio-implant with 20 ml of D5LR.
11. Fully submerge the bio-implant for 10 minutes and then gently flush the bio-implant again with 20 ml of D5LR.
12. Once dilution is completed, inspect bio-implant integrity to ensure no damage occurred during the thawing and dilution process.
13. Perform leak testing under pressure prior to implanting.
 - a. Inspect the ties for any loosening.
 - b. Perform additional ligations or reinforcements as needed.
14. Remove the cannulae from the allograft.



When performing pressurization of the bio-implant to check integrity or valvular competency, do not over-pressurize. This could damage the bio-implant.

FOR SAPHENOUS AND FEMORAL IMPLANTS:

(Cont. from Opening and Thawing Instructions)

8. **Non-Sterile Team Member:** Pour 1000 ml of room temperature D5LR into the second sterile major basin.
9. **Sterile Team Member:** Gently remove the bio-implant from the pouch with smooth forceps and carefully place into the basin containing the D5LR. Cut the non-cannulated end of the vessel at suture tie to allow perfusion.
10. Using the 20 cc syringe, gently flush the bio-implant with 20 ml of D5LR.
11. Fully submerge the bio-implant for 10 minutes and then gently flush the bio-implant again with 20 ml of D5LR.
12. Once dilution is completed, inspect bio-implant integrity to ensure no damage occurred during the thawing and dilution process.
13. Perform leak testing under pressure prior to implanting.
 - a. Inspect the ties for any loosening.
 - b. Perform additional ligations or reinforcements as needed.
14. Remove the cannula from the allograft.

FOR AORTOILIAC ARTERY, DISTAL AORTOILIAC, ABDOMINAL AORTA BIO-IMPLANTS:

(Cont. from Opening and Thawing Instructions)

8. **Non-Sterile Team Member:** Pour 1000 mL of room temperature D5LR into the second sterile major basin.
9. **Sterile Team Member:** Gently remove the bio-implant from the pouch with smooth forceps, allow cryosolution to drain completely into the first major basin and carefully place into the second major basin containing the D5LR.
10. **Sterile Team Member:** Fully submerge the bio-implant to remove all air bubbles and allow the bio-implant to passively rinse fully submerged in the solution for a minimum of 5 minutes.
11. Once dilution is completed, inspect bio-implant integrity to ensure no damage occurred during the thawing and dilution process.
12. Perform leak testing under pressure prior to implanting.
 - a. Inspect the ties for any loosening.
 - b. Perform additional ligations or reinforcements as needed.
13. It is recommended to position the bio-implant with lumbar artery ligations facing anterior to more easily facilitate evaluation of hemostasis.



Do not allow the bio-implant to dry. If implantation is delayed, cover basin with sterile towel and place basin in larger basin of sterile slush.